

Guide to using Tristel Duo on TV Ultrasound Probes

CLEANING



Clean the probe with a Clinell Wipe

DISINFECTION



Dispense ONE aliquots of Tristel Duo ULT onto a clean paper towel.



Use the towel to spread the foam over the surface of the probe to cover all areas.

DRYING



Leave the surface to dry to ensure a minimum 30 seconds contact time.